§890.3920

the current good manufacturing practice requirements of the quality system regulation in part 820 of this chapter, with the exception of §820.180, regarding general requirements concerning records, and §820.198, regarding complaint files.

 $[61\ FR\ 8439,\ Mar.\ 4,\ 1996,\ as\ amended\ at\ 66\ FR\ 38817,\ July\ 25,\ 2001]$

§890.3920 Wheelchair component.

- (a) Identification. A wheelchair component is a device intended for medical purposes that is generally sold as an integral part of a wheelchair, but may also be sold separately as a replacement part. Examples of wheelchair components are the following: Armrest, narrowing attachment, belt, extension brake, curb climber, cushion, antitip device, footrest, handrim, hill holder, leg rest, heel loops, and toe loops.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §890.9.
- [48 FR 53047, Nov. 23, 1983, as amended at 59 FR 63014, Dec. 7, 1994; 66 FR 38817, July 25, 2001]

§890.3930 Wheelchair elevator.

- (a) Permanently mounted wheelchair platform lift—(1) *Identification*. A permanently mounted wheelchair platform lift is a motorized vertical or inclined platform lift device permanently installed in one location that is intended for use in mitigating mobility impairment caused by injury or other disease by providing a guided platform to move a person from one level to another, with or without a wheelchair.
- (2) Classification. Class II. The permanently mounted wheelchair platform lift is exempt from premarket notification procedures in subpart E of part 807 of this chapter, subject to §890.9 and the following conditions for exemption:
- (i) Appropriate analysis and nonclinical testing (such as that outlined in the currently FDA-recognized edition of ASME A18.1 "Safety Standard for Platform Lifts and Stairway Chair Lifts") must demonstrate that the safety controls are adequate to prevent

a free fall of the platform in the event of a device failure;

- (ii) Appropriate analysis and nonclinical testing (such as that outlined in the currently FDA-recognized edition of ASME A18.1 "Safety Standard for Platform Lifts and Stairway Chair Lifts") must demonstrate the ability of the device to withstand the rated load with an appropriate factor of safety;
- (iii) Appropriate analysis and nonclinical testing (such as that outlined in the currently FDA-recognized edition of ASME A18.1 "Safety Standard for Platform Lifts and Stairway Chair Lifts") must demonstrate the ability of the enclosures to prevent the user from falling from the device; and
- (iv) Appropriate analysis and nonclinical testing (such as that outlined in the currently FDA-recognized editions of AAMI/ANSI/IEC 60601-1-2, "Medical Electrical Equipment—Part 1-2: General Requirements for Safety— Collateral Standard: Electromagnetic Compatibility—Requirements and Tests," and ASME A18.1 "Safety Standard for Platform Lifts and Stairway Chair Lifts") must validate electromagnetic compatibility and electrical safety.
- (b) Portable wheelchair elevators—(1) Identification. A portable wheelchair elevator is a motorized lift device that is not permanently mounted in one location and that is intended for use in mitigating mobility impairment caused by injury or other disease by providing a means to move a person, with or without a wheelchair, from one level to another (e.g., portable platform lifts, attendant-operated stair climbing devices for wheelchairs).
 - (2) Classification. Class II.

[78 FR 14015, Mar. 4, 2013]

§890.3940 Wheelchair platform scale.

- (a) *Identification*. A wheelchair platform scale is a device with a base designed to accommodate a wheelchair. It is intended for medical purposes to weigh a person who is confined to a wheelchair.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §890.9. The device is also exempt from the current

Food and Drug Administration, HHS

good manufacturing practice requirements of the quality system regulation in part 820 of this chapter, with the exception of §820.180, regarding general requirements concerning records and §820.198, regarding complaint files.

[48 FR 53047, Nov. 23, 1983, as amended at 59 FR 63015, Dec. 7, 1994; 66 FR 38817, July 25, 2001]

Subpart E [Reserved]

Subpart F—Physical Medicine Therapeutic Devices

§890.5050 Daily activity assist device.

- (a) *Identification*. A daily activity assist device is a modified adaptor or utensil (e.g., a dressing, grooming, recreational activity, transfer, eating, or homemaking aid) that is intended for medical purposes to assist a patient to perform a specific function.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §890.9. If the device is not labeled or otherwise represented as sterile, the device is also exempt from the current good manufacturing practice requirements of the quality system regulation in part 820 of this chapter, with the exception of §820.180, regarding general requirements concerning records and §820.198, regarding complaint files.

 $[48\ FR\ 53047,\ Nov.\ 23,\ 1983,\ as\ amended\ at\ 66\ FR\ 38817,\ July\ 25,\ 2001]$

\$890.5100 Immersion hydrobath.

- (a) Identification. An immersion hydrobath is a device intended for medical purposes that consists of water agitators and that may include a tub to be filled with water. The water temperature may be measured by a gauge. It is used in hydrotherapy to relieve pain and itching and as an aid in the healing process of inflamed and traumatized tissue, and it serves as a setting for removal of contaminated tissue.
- (b) Classification. Class II (performance standards).

§890.5110 Paraffin bath.

- (a) Identification. A paraffin bath is a device intended for medical purposes that consists of a tub to be filled with liquid paraffin (wax) and maintained at an elevated temperature in which the patient's appendages (e.g., hands or fingers) are placed to relieve pain and stiffness.
- (b) Classification. Class II (performance standards).

§890.5125 Nonpowered sitz bath.

- (a) Identification. A nonpowered sitz bath is a device intended for medical purposes that consists of a tub to be filled with water for use in external hydrotherapy to relieve pain or pruritis and to accelerate the healing of inflamed or traumatized tissues of the perianal and perineal areas.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §890.9. The device is also exempt from the current good manufacturing practice requirements of the quality system regulation in part 820 of this chapter, with the exception of §820.180, regarding general requirements concerning records and §820.198, regarding complaint files.

[48 FR 53047, Nov. 23, 1983, as amended at 54 FR 25052, June 12, 1989; 66 FR 38818, July 25, 2001]

§890.5150 Powered patient transport.

- (a) Powered patient stairway chair lifts—(1) *Identification*. A powered patient stairway chair lift is a motorized lift equipped with a seat and permanently mounted in one location that is intended for use in mitigating mobility impairment caused by injury or other disease by moving a person up and down a stairway.
- (2) Classification. Class II. The stairway chair lift is exempt from premarket notification procedures in subpart E of part 807 of this chapter, subject to §890.9 and the following conditions for exemption:
- (i) Appropriate analysis and nonclinical testing (such as that outlined in the currently FDA-recognized edition of American Society of Mechanical Engineers (ASME) A18.1 "Safety Standard